

FOOTMOTIONS.



STAND-ALONE
SCREWS



FOOTMOTION S

Indication for use: The implants of the Foot and Hand Motion range are intended for the fixation of bone fractures and osteotomies and for arthrodeses of foot and hand in adults.



Contraindications :

- Pregnancy.
- Acute or chronic local or systemic infections.
- Allergy to one of the materials used or sensitivity to foreign bodies.

A COMPREHENSIVE RANGE OF SOLUTIONS FOR FOREFOOT TREATMENT


→ SCREW FEATURES

- Ø2.25, Ø2.6, Ø3.0 mm cannulated screws
- Self-drilling, self-tapping and self-compressive screws
- Hexalobular screw recess design
- Compression with long distal thread feature



→ INSTRUMENTATION FEATURES

- Instrument mutualisation for Ø2.6 & Ø3.0 mm screws: same pin, drill, countersink and screwdriver
- Patented 3 in 1 instrument for a simple use: differential gauge for pin, countersink and screwdriver
- Compatible with a standard quick coupling system




FOOTMOTION S

MINIMALLY INVASIVE SURGERY WITH Ø3.5 MM CHAMFERED HEAD SCREWS

→ SCREW FEATURES

- Ø3.5 mm cannulated screws
- Hexalobular screw recess design
- Self-drilling and self-tapping screws
- Positioning screw with full thread



→ INSTRUMENTATION FEATURES

- Patented 3 in 1 instrument for a simple use: differential gauge for pin, countersink and screwdriver
- Compatible with a standard quick coupling system
- Head profile marking on the patented 3 in 1 instrument to help visualize screw head orientation




SNAP-OFF SCREWS AND INSTRUMENT

→ IMPLANT

➤ Ø2.0 mm snap-off screw dedicated to Weil osteotomy, from L11 to L15 mm (1 mm increments)

- Self-drilling and self-tapping extremity ensuring the penetration in cortical bone (a)
- Self-drilling of the screw head allowing a partial burying to limit soft tissue irritation (b)
- Snap-off area buried under the screw head
- Compatible with wire driver (Ø2.4 mm)





Snap-off screw : the screw separates itself from the shank as soon as the head gets in contact with the bone.

N.B: For an additional compression, use the screwdriver (ANC770).

→ DEDICATED INSTRUMENT: SCREWDRIVER

➤ Screwdriver allowing:

- The separation between the shank and the screw
- The finalization of the compression after separation
- The implant removal performance



ANC770


SURGICAL TECHNIQUE

CHEVRON OSTEOTOMY

Example of a Chevron osteotomy surgical technique using a Ø2.6 mm self-compressive screw (H1.1HFT2.6LxxD)



1. Insert the Ø1.0 L90 mm pin (ANC1451) until the opposite cortex is reached.




2. Slide the 3 in 1 instrument (ANC1282) along the pin to the bone (a). Read directly the length of the screw on the measuring gauge part of the 3 in 1 instrument at the rear of the pin (b).


N.B: At this stage the pin can be inserted deeper in order to prevent its removal during drilling step.



3. To ease the screw insertion, countersink the first cortex using the 3 in 1 instrument (ANC1282).



4. Drill to the desired depth with the Ø2.0 mm quick coupling drill bit (ANC1341). The drilling depth can be checked using the markings on the drill (c).



5. Select the appropriate screw length and insert the self-compressive screw along the pin using the screwdriver part of the pin using the screwdriver part of the 3 in 1 instrument (ANC1282) until the desired reduction and compression are achieved and the screw head is buried. Then remove the pin.




FINAL RESULT


SURGICAL TECHNIQUE

MINIMALLY INVASIVE OSTEOTOMY

Example of a minimally invasive osteotomy surgical technique using a Ø3.5 mm chamfered head screw (H1.5NT3.5LxxD)




1. Insert the Ø1.4 mm pin (ANC1657) to maintain the correction.




2. Slide the 3 in 1 instrument (ANC1286) along the pin until the cortex is reached (a). With the measuring gauge part of this instrument, read the insertion depth using the marking on the pin (b).


N.B: At this stage, the pin can be inserted deeper in order to prevent its removal during the drilling step.



3. Drill to the desired depth with the Ø2.5 mm quick coupling drill bit (ANC1718). The drilling depth can be checked using the markings on the drill (c).




4. To ease the screw insertion, countersink the first cortex using the 3 in 1 instrument (ANC1286).



5. Select the appropriate screw length and insert the chamfered head screw along the pin using the screwdriver part of the 3 in 1 instrument (ANC1286) until the desired reduction is achieved.


Check with fluoroscopy that the chamfered head is flush to the bone and then remove the pin.

FINAL RESULT



Repeat the same steps for the insertion of the second screw.


N.B: This surgical technique can be associated with the use of hallux valgus targeting guide. Please refer to the G-HALVA brochure for more information.



SURGICAL TECHNIQUE

WEIL OSTEOTOMY


Example of a Weil osteotomy surgical technique using a Ø2.0 mm snap-off Weil screw (WST2.0Lxx-ST)



1. Perform a first horizontal cut using the oscillating saw starting at the junction of the dorsal cartilage (a). Then, perform the second parallel dorsal cut with the first cut and then remove the bone fragment (b).



2. The reduction is made manually by flexing the toe.



3. Insert the screw with the power tool. As soon as the compression is finished, the screw snaps off. If necessary, use the screwdriver (ANC770) with the handle (ANC350) to make the compression.

N.B: In case of a hard cortical bone it is recommended to prepare the screw insertion using a pin.

Caution: In osteoporotic bone, it may be necessary to provoke release of the shank prior screw head reaches the cortical bone to avoid excessive screwing.




FINAL RESULT

IMPLANTS REFERENCES


→ FOREFOOT SCREWS

Ø2.25 MM CANNULATED SCREWS

Ref	Description	
H0.9HFT2.25LxxD*	Ø2.25 mm self-compressive screw - cannula Ø0.9 - long thread - L10 mm to 34 mm (2 mm increments)	


*Non anodized

Ø2.6 MM CANNULATED SCREWS

Ref	Description	
H1.1HFT2.6LxxD*	Ø2.6 mm self-compressive screw - cannula Ø1.1 - long thread - L10 mm to 34 mm (2 mm increments)	


*Yellow anodized

Ø3.0 MM CANNULATED SCREWS

Ref	Description	
H1.1HFT3.0LxxD*	Ø3.0 mm self-compressive screw - cannula Ø1.1 - long thread - L10 mm to 40 mm (2 mm increments)	

*Fuchsia anodized


Ø3.5 MM CANNULATED SCREWS

Ref	Description	
H1.5NT3.5LxxD*	Ø3.5 mm chamfered head screw - cannula Ø1.5 - L20 mm to 60 mm (2 mm increments)	

*Dark blue anodized

→ SNAP-OFF WEIL SCREWS

Ø2.0 MM SNAP-OFF SCREWS

Ref	Description	
WST2.0Lxx-ST*	Ø2.0 mm snap-off Weil screw - STERILE - L11 mm to 15 mm (1 mm increments)	

*Non anodized

Remark: please note that the «xx» in the references represents the length of the screw.
Ex: the reference for the Ø2.6 mm self-compressive screw - cannula Ø1.1 - long thread - L30 mm is «H1.1HFT2.6L30D»

Remark: please note that all implants are also available in sterile packaging. An 'ST' code is added at the end of the reference.
Ex : « H1.1HFT2.6L30D-ST »

INSTRUMENTS REFERENCES

INSTRUMENTS FOR Ø2.25 MM SCREWS*

Ref.	Description	Qty
ANC350	Ø4.5 mm AO quick coupling handle - Size 1	1
ANC770	Screwdriver for Ø2.0 mm Snap-off screws	1
ANC1280	3 in 1 instrument for Ø2.25 mm screws	1
ANC1446	Ø1.7 mm quick coupling drill bit - cannula Ø0.9 mm - L75 mm	1
ANC1450	Pin Ø0.8 L90 mm	6

*Snap-off screwdriver is available in this kit

OPTIONAL INSTRUMENTATION

Ref.	Description
ANC144	16 cm forceps
ANC1446-ST	Ø1.7 mm quick coupling drill bit - cannula Ø0.9 mm - L75 mm - STERILE
ANC1450-ST	Pin Ø0.8 L90 mm – STERILE
ANC1672	T7 prehensor screwdriver with AO quick coupling system - cannula Ø0.9 mm

INSTRUMENTS FOR Ø2.6 & Ø3.0 MM SCREWS*

Ref.	Description	Qty
ANC350	Ø4.5 mm AO quick coupling handle - Size 1	1
ANC770	Screwdriver for Ø2.0 mm Snap-off screws	1
ANC1282	3 in 1 instrument for Ø2.6 & Ø3.0 mm screws	1
ANC1341	Ø2.0 mm quick coupling drill bit - cannula Ø1.1 mm - L 80 mm	1
ANC1451	Pin Ø1.0 L90 mm	6

*Snap-off screwdriver is available in this kit

OPTIONAL INSTRUMENTATION

Ref.	Description
ANC144	16 cm forceps
ANC1341-ST	Ø2.0 mm quick coupling drill bit - cannula Ø1.1 mm - L 80 mm - STERILE
ANC1451-ST	Pin Ø1.0 L90 mm – STERILE
ANC1673	T8 prehensor screwdriver with AO quick coupling system - cannula Ø1.1 mm

INSTRUMENTS FOR Ø3.5 MM SCREWS

Ref.	Description	Qty
ANC350	Ø4.5 mm AO quick coupling handle - Size 1	1
ANC1286	3 in 1 instrument for Ø3.5 mm screws	1
ANC1657	Pin Ø1.4 L150 mm	6
ANC1674	T10 prehensor screwdriver with AO quick coupling system - cannula Ø1.5 mm	1
ANC1718	Ø2.5 mm quick coupling drill bit - cannula Ø1.5 mm - L 110 mm	1

OPTIONAL INSTRUMENTATION

Ref.	Description
ANC1657-ST	Pin Ø1.4 L150 mm – STERILE
ANC1718-ST	Ø2.5 mm quick coupling drill bit - cannula Ø1.5 mm - L 110 mm - STERILE

N.B: Diameter of screw(s) to be used with each 3 in 1 instrument is marked on the instruments.

REMOVAL KIT

If you have to remove Footmotion S implants, make sure to order the **Newclip Technics** removal set, which includes the following instruments:

- ANC350: Ø4.5 mm AO quick coupling handle - Size 1
- ANC770: Screwdriver for Ø2.0 mm Snap-off screws
- ANC1672: T7 prehensor screwdriver with AO quick coupling system - cannula Ø0.9 mm
- ANC1673: T8 prehensor screwdriver with AO quick coupling system - cannula Ø1.1 mm
- ANC1674: T10 prehensor screwdriver with AO quick coupling system - cannula Ø1.5 mm
- ANC1450: Pin Ø0.8 L90 mm
- ANC1451: Pin Ø1.0 L90 mm
- ANC1657: Pin Ø1.4 L150 mm

MIS INSTRUMENTS


Ref.	Description	Qty
ANC602	Elevator for percutaneous surgery	1
ANC606	Elevator - rugine for percutaneous surgery	1
6051*	Sawblade handle	1
GS 61.6380**	Rasp 2 mm angled for percutaneous surgery	1

*Manufacturer: SA SWANN MORTON SINNER
Medical Class: I


**Manufacturer: GSOURCE
Medical Class: I

KIT DESCRIPTION

KIT FOR Ø2.25 MM SCREWS




KIT FOR Ø2.6 & Ø3.0 MM SCREWS




KIT DESCRIPTION

KIT FOR Ø3.5 MM SCREWS




KIT FOR MIS INSTRUMENTS




CONFIGURATIONS

- Additional larger boxes are available to group multiple kits together.


BOX FOR 2 KITS



BOX FOR 3 KITS



BOX FOR 4 KITS



This information is intended to demonstrate the Newclip Technics portfolio of medical devices. Always refer to the package insert, product label and/or user instructions before using any Newclip Technics product. These products must be handled and/or implanted by trained and qualified staff who have read the instructions before use. A surgeon must always rely on her or his own professional clinical judgement when deciding whether to use a particular product when treating a particular patient. Product availability is subject to the regulatory or medical practices that govern individual markets. Please contact your Newclip Technics representative if you have questions about the availability of Newclip Technics products in your area.

Manufacturer: Newclip Technics - Brochure US – Footmotion S – ED2 – 04/2025 - Medical device - US Class: II - Read labelling and instructions before the use of Newclip Technics medical devices. These products must be handled and/or implanted by trained and qualified staff who have read the instructions before use. Non-contractual pictures.

Newclip Technics - 45 rue des Garotières - 44115 Haute Goulaine, France. Our subsidiaries: Newclip USA - Newclip Australia - Newclip GMBH - Newclip Japan - Newclip Iberia - Newclip Belgium.

newcliptechnics.com

